



LINGUIST CERTIFICATION

This Translation Certificate confirms the included documents have been completed and are, to the best knowledge and belief of the below linguist engaged on the project, full and accurate translations of the source material.

File Name: KN95 Chunyu report verification

Source Language: Chinese

Target Language: English

Linguist Name: Michael Day

TransPerfect Office Partnership: TransPerfect London

Office Location: 1st Floor, 33 Aldgate High Street, London, EC3N 1AH

List of Linguist Qualifications:

MA East Asian Languages and Cultures

Linguist Signature:


Michael Day (Apr 29, 2020)

Name: Michael Day

Title: Translator

Date: 4/29/2020

Authorized Signature:


Lauren Sicurella (Apr 29, 2020)

Name: Lauren Sicurella

Title: Senior Project Manager

Date: 4/29/2020

Reason for signature: I approve the accuracy of this document content as written

Following a certified translation of the original test report. Chinese Test Report can be verified online, accessing to <https://gqi.org.cn/gqicqs/jsp/web/eng/checkquery/NonCheckQuery.xhtml> and using the Test Report number FZ2009902, and the Random Number HWM6826

No. FZ2009902

[logo:]
MA
201719002006

[logo:]
ilac-MRA

[logo:]
CNAS

Chinese
National
Accreditation
Service for
Conformity
Assessment
Testing TESTI
NG CNAS
L0153

[logo:] AL
2017 (Guangdong) Quality
Control Approval Number
039

T e s t R e p o r t

TEST REPORT

Client	Shandong Chunyujinxiu Textiles Co., Ltd.
Manufacturer	Shandong Chunyujinxiu Textiles Co., Ltd.
Sample name	KN95 protective mask
Model number specification	KN95
Test category	Contracted test



**Guangdong Testing Institute of Product Quality
Supervision**

[logo]

GUANGDONG TESTING INSTITUTE OF PRODUCT QUALITY SUPERVISION

02 of 07 |

Guangdong Testing Institute of Product Quality Supervision

GUANGDONG TESTING INSTITUTE OF PRODUCT QUALITY SUPERVISION
[barcode]

Test Report**TEST REPORT**

Random report number: HWM6826

Page 1/5

Sample name	KN95 protective mask		Sample number	YFZ20/009902
	Send sample (√)	Select sample (/)		
Trademark	---		Model number specification	KN95
Client	Shandong Chunyujinxiu Textiles Co., Ltd.		Test category	Contracted test
Client address	98 Shengtai Rd, Zhaoyuan Shi, Yantai Shi, Shandong Sheng		Product number/lot number	---
Manufacturer	Shandong Chunyujinxiu Textiles Co., Ltd.		Sample selection number	---
Company tested	---		Date of manufacture	---
Sampling company	---		Sample quantity	40 (pcs.)
Sampling place	---		Sampling base number	---
Sampling date	---		Examination place	Our laboratory
Sample receipt date	April 7, 2020		Examination date	April 7, 2020 - April 15, 2020
Basis for test	GB 2626-2006 <Breathing protection device: Self-breathing filter particle protection respirator>			
Basis for determination	---			
Testing determination	<p>This contracted test examined a total of seven items, and the items examined met the requirements of the standards.</p> <p>[e-signature:] Guangdong Testing Institute of Product Quality Supervision Special examination and testing seal (Q1) (Special examination and testing seal) Signing and issuance date: April 15, 2020</p>			
Notes	In the report, "---" indicates that an item is not applicable, and "/" indicates that the item is blank.			

Approval: [signature]

Examination [signature] and verification:

Lead examiner: [signature]

Examination Report

TEST REPORT

[e-signature:] Guangdong Testing Institute of
Product Quality Supervision
Special examination and testing seal
(Q1)

Serial number	Test item [company]	Standard Provision	Standard requirement	Test result		Single item Conclusion	Notes
1	Filter efficiency [%]	5.3	KN95 \geq 95.0	Not pretreated	99.5	Approval	/
					99.6		
					99.2		
					99.5		
					99.5		
					99.5		
					99.4		
					99.6		
					99.4		
					99.5		
			Pretreated	99.6			
				99.5			
				99.6			
				99.5			
Sodium chloride particle testing temperature: (25 \pm 5) °C humidity: (30 \pm 10) %	Test temperature: (25) °C test temperature: (30) %	99.6					
		99.6					
2	Inspiratory resistance [Pa]	5.5	Total inspiratory resistance \leq 350	Not pretreated	196.6	Approval	/
					193.7		
				Pretreated	204.1		
					200.7		
3	Expiratory resistance [Pa]	5.5	Total expiratory resistance \leq 250	Not pretreated	97.6	Approval	/
					95.4		
				Pretreated	96.5		
					97.1		

Examination Report

TEST REPORT

[e-signature:] Guangdong Testing Institute of
Product Quality Supervision
Special examination and testing seal
(Q1)

Serial number	Test item [company]	Standard Provision	Standard requirement	Test result		Single item Conclusion	Notes
4	Dead space [%]	5.7	When indicated in terms of proportion of carbon dioxide inhaled by volume, average value of results should be ≤ 1	Average value: 1.0		Approval	/
5	Headband	5.9	Each headband, buckle and other adjustment part of the disposable face mask should endure stretching of 10 N for a duration of 10 s without slipping or breaking	Not pretreated	No slipping or breakage occurred	Approval	/
				Pretreated	No slipping or breakage occurred		
6	Connection and connecting parts	5.10	Under the stipulated testing conditions for replaceable half face masks, all connections and connecting parts between the replaceable filter component and the mask should endure stretching of 50 N for a duration of 10 s without slipping, breaking, or deforming	Not pretreated	No connecting parts, this item not tested	---	/
				Pretreated			
7	Flammability	5.13	No part should catch fire after exposed to flame from the time the flame is removed; if it does catch fire, it should not burn continuously for more than 5s	Not pretreated	No burning occurred	Approval	/
				Pretreated	No burning occurred		

Examination Report

TEST REPORT

[e-signature:] Guangdong Testing Institute of
Product Quality Supervision
Special examination and testing seal
(Q1)

Photograph and
explanation



Sample description

/

Model specification or
other explanation

KN95

Examination Report

TEST REPORT

[e-signature:] Guangdong Testing Institute of
Product Quality Supervision
Special examination and testing seal
(Q1)

Additional notes:

1. Examination place: 10 Kexue Dadao, Kexue Cheng, Huangpu Qu, Guangzhou Shi
2. Address and postal code of client: 98 Shengtai Rd, Zhaoyuan Shi, Yantai Shi, Shandong Sheng
3. Examination environment conditions: All test items carried out under the conditions stipulated by the applicable standards (except as noted)
4. Sample selection process (if applicable): -----
5. Explanation of method deviating from standard (if applicable):-----
6. Explanation of degree of uncertainty of test results (if applicable):-----
7. Subcontractor of subcontracted items (if applicable):-----

FDA's Establishment Registration & Device Listing |

(FDA's Verification database: <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/rl.cfm> .
Set Owner/Operator number to 10064085, then click on search)

New Search	Back To Search Results
<p>Establishment: SHANDONG CHUNYU JINXIU TEXTILES CO., LTD. No. 98 Shengtai Road, Zhaoyuan City Yantai Shandong, CN 265400 Status: Active; Awaiting Assignment Of Registration Number Date Of Registration Status: 2020</p> <p>Owner/Operator: Shandong Chunyu Jinxiu Textiles Co., Ltd. No. 98 Shengtai Road, Zhaoyuan City Yantai, Shandong CN 265400 Owner/Operator Number: 10064085</p> <p>Official Correspondent: Amy Zhang No. 98 Shengtai Road, Zhaoyuan City Yantai, Shandong CN 265400 Phone: 86-0535-8154001</p> <p>US Agent: Aizhen Chen 143 Division St New York, NY US 10002 Phone: 917 2089162 Ext Email: 19757348@Qq.Com</p>	
<p>* Firm Establishment Identifier (FEI) should be used for identification of entities within the imports message set</p>	

New Search	Back To Search Results
<p>Proprietary Name: Mask: CY20-KZ001, CY20-KZ002 Classification Name: FACE MASK (EXCEPT N95 RESPIRATOR) FOR GENERAL PUBLIC/HEALTHCARE PERSONNEL PER IIE GUIDANCE Product Code: QKR Device Class: Not Classified Registered Establishment Name: SHANDONG CHUNYU JINXIU TEXTILES CO., LTD. Owner/Operator: Shandong Chunyu Jinxiu Textiles Co., Ltd. Owner/Operator Number: 10064085 Establishment Operations: Foreign Exporter; Manufacturer</p>	

